

Certificate of Analysis

Batch Number: B 534

General

This product is non-hazardous. Safety:

2-8°C. Do not freeze. Storage:

ESCG100 **Catalogue Number:**

Preparation Date: 26 March 2015 24 July 2015 **Expiration Date:**

1.215 ml +/-32 ul Volume: Buffered saline solution. Suspension media:

Sterilisation method: Gamma irradiation.

Counts

(Method Ref: ESCG-001)

Mean

St.Dev. 1.4

Cryptosporidium count:

99

St.Dev.

Giardia count:

Mean 100

1.8

DAPI staining:

Cryptosporidium % +ve

100

%

Giardia % +ve

100 %

Stock specifics

Organism: Cryptosporidium parvum

Strain: Iowa Source: **Bovine**

Shed date: 06 March 2015

Purification method: Discontinuous sucrose and cesium chloride centrifugation gradients.

Organism: Giardia lamblia

Strain: H3

Source: Gerbil

Shed date: 12 March 2015

Purification method: Sucrose and Percoll density gradient centrifugation

Quality Officer:

Linh Le

TECHNICAL

This document is issued in accordance with NATA's accreditation requirements.
Accredited for compliance with ISO Guide 34.

Accredited Reference Material Producer Accreditation Number: 14993 Site Number: 15276

Water Industry Enquires:

BTF Pty Ltd. Sydney Australia Tel. 61 (0) 2 8877 9150 Fax. 61 (0) 2 8877 9101 www.btfbio.com



