

## Certificate of Analysis

Batch Number: B 629

## General

**Safety:** This product is non-hazardous.

Storage: 2-8°C. Do not freeze.

Catalogue Number: ESCG100-5
Preparation Date: 28 March 2018
Expiration Date: 26 July 2018

Volume: 1.303 ml +/- 32 ul Suspension media: Buffered saline solution.

Sterilisation method: Gamma irradiation.

**Counts** 

(Method Ref: ESCG-001)

Mean

St.Dev.

Cryptosporidium count:

100

**2.2** St.Dev.

Giardia count:

Giardia % +ve

Mean **100** 

2.4

**DAPI staining:** 

Cryptosporidium % +ve

100

100

%

## Stock specifics

Organism: Cryptosporidium parvum

Strain: Iowa
Source: Bovine

Shed date: 21 March 2018

Purification method: Discontinuous sucrose and cesium chloride centrifugation gradients.

Organism: Giardia lamblia

Strain: H3

Source: Gerbil

Shed date: 19 March 2018

Purification method: Sucrose and Percoll density gradient centrifugation

**Quality Officer:** 

lihle

Linh Le

NATA
ACCREDITED FOR
TECHNICAL

This document is issued in accordance with NATA's accreditation requirements.
Accredited for compliance with ISO Guide 34.

Accredited Reference Material Producer Accreditation Number: 14993 Site Number: 15276

Water Industry Enquires:

BTF Pty Ltd. Sydney Australia Tel. 61 (0) 2 8877 9150 Fax. 61 (0) 2 8877 9101 www.btfbio.com



