

Batch Number: **B 104****Fluorescence analysis - Mean Fluorescence Intensity***Cryptosporidium* oocysts: 937.86*Giardia* cysts: 931.20**Non-specific binding assay**

Number of particles detected: 0

Control number of particles detected: 3

**General**

Safety: This product contains sodium azide. Handle with care.

Storage: 2-8°C. Do not freeze.

Catalogue Number: EST-CG20

Preparation Date: 11 June 2014

Expiration Date: 12 June 2015

**Quality Officer:**

Linh Le

**Water Industry Enquires:**

BTF Pty Ltd.  
Sydney Australia  
Tel. 61 (0) 2 8877 9150  
Fax. 61 (0) 2 8877 9101  
[www.btfbio.com](http://www.btfbio.com)

