

Batch Number: **B 104****Fluorescence analysis - Mean Fluorescence Intensity***Cryptosporidium* oocysts: 937.86*Giardia* cysts: 931.20**Non-specific binding assay**

Number of particles detected: 0

Control number of particles detected: 3

General

Safety: This product contains sodium azide. Handle with care.
Storage: 2-8°C. Do not freeze.
Catalogue Number: EST-CG80
Preparation Date: 11 June 2014
Expiration Date: 12 June 2015

Quality Officer:

Linh Le

**Water Industry Enquires:**

BTF Pty Ltd.
Sydney Australia
Tel. 61 (0) 2 8877 9150
Fax. 61 (0) 2 8877 9101
www.btfbio.com

