

Batch Number: **B 108**

**Fluorescence analysis - Mean Fluorescence Intensity**

*Cryptosporidium* oocysts: 1107.75

**Non-specific binding assay**

Number of particles detected: 1

Control number of particles detected: 2

**General**

Safety: This product contains sodium azide. Handle with care.  
Storage: 2-8°C. Do not freeze.  
Catalogue Number: EST-C400  
Preparation Date: 27 November 2014  
Expiration Date: 27 November 2016

**Quality Officer:**



.....  
Linh Le



**Water Industry Enquires:**

BTF Pty Ltd.  
Sydney Australia  
Tel. 61 (0) 2 8877 9150  
Fax. 61 (0) 2 8877 9101  
[www.btfbio.com](http://www.btfbio.com)

